



Notified Body 0321

Issued to: Ultimate Cleaners Industrial Ltd
Victoria House
Colliery Road
WOLVERHAMPTON
West Midlands, WV1 2RD
UK

SATRA Client : P9829

EC Type-Examination Certificate

Number 5812 Issue 3

Date first issued: 30/07/2012

This is to certify that the product group reference "Group TCG" comprising the following products:

Product Reference	Description
Group TCG	Five fingered 32 oz. Terry cloth gloves and sleeves
TCG32K	With knitted elasticated wrist
TCG32DXX	With canvas cuff (XX suffix indicates cuff length between 4-12 inches)
TCG32TXX	With terry cuff (XX suffix indicates cuff length between 4-12 inches)
TCMG1	Mitten Variant with 8inch terry cuff
TCGSLXX	Sleeves as group above with elasticated knitted wrist and elasticated top (XX suffix indicates sleeve length between 12-20 inches)

Classification:

EN 388:2016	Level	EN407:2004	Level
Abrasion resistance	2	Burning Behaviour	X
Blade cut resistance	2	Contact Heat	2
Tear resistance	3	Convective Heat	X
Puncture resistance	2	Radiant Heat	X
TDM Cut Resistance	X	Small Splashes of Molten Metal	X
		Large Quantities of Molten Metal	X

Technical reports:

SPC0203477/1217/1, SPC0203477/1217/MB, SPC0258435/1723/7, SPC0259817/1728/3

has been subject to an EC Type-examination in accordance with Article 10 of the PPE Directive (89/686/EEC) and has been shown to satisfy the relevant provisions of this Directive for the intermediate category through:

- i Testing to the following standard: **EN388:2016; EN420:2003+A1:2009; EN407: 2004**
- ii Examination of the relevant technical documentation.

You are therefore licensed to mark the product(s) listed above in accordance with Article 13 of Directive (89/686/EEC) and any relevant amending Directives once you have drawn up an EC declaration of product conformity. Please note that:

1. Full details of the certification and product are contained in the manufacturer's technical file
2. This certificate is issued subject to the conditions on the reverse side of this certificate
3. Where a translation of this certificate exists, the English language version shall be considered as the authoritative text

Signed:

(Andrew Craggs)

Date 21/08/2017

Signed:

On behalf of SATRA

CONDITIONS AND NOTES

1. CARE AND PROTECTION OF CERTIFICATES

This certificate shall be kept together with the relevant technical files in a safe place by the client named on this certificate. Production of this certificate and other documentation may be required by a representative of the EC member state government (in the UK, this will be Trading Standards Department of Local Authorities) and these must be produced within the specified time.

2. EXPIRY OF CERTIFICATES

This certificate relates only to the condition of the testable items at the time of the certification procedure. This certificate will cover the product until such a time that any of these conditions are changed. The certificate therefore does not have an expiry date.

3. WITHDRAWAL OF CERTIFICATES

SATRA Technology Centre reserves the right to withdraw this certificate if it is found that a condition of manufacture, design, materials or packaging have been changed and therefore no longer comply with the requirements of the PPE directive.

4. LOSS OF CERTIFICATES

If you lose this certificate you can get a duplicate from SATRA Technology Centre for a minimal charge by quoting details such as the approximate date of issue and the product reference.

PPE Doc 13/14 Issue 1